



QPS Evaluation Services Inc
Testing, Certification and Field Evaluation Body
Accredited in Canada, the USA, and Internationally

File
LR1197

CERTIFICATE OF COMPLIANCE
(ISO TYPE 3 CERTIFICATION SYSTEM)

Issued to	Orion Concepts Inc.
Address	2-451 Enfield Road Burlington, ON L7T 2X5 Canada
Project Number	LR1197-1R5
Product	LED Laryngoscope Handle; LED Handle for Sigmoidoscope/Anoscope Illuminators
Model Numbers	GS-MED-A-XX, GS-STU-A-NU and GS-PEN-A-XX, where XX may be either NA or EU or NU or blank or Other and represents different mechanical hook/latch configurations only; Model GS-LEI-X-A-HP, where X may be S or blank or other
Ratings	3.0Vdc, internally powered
Applicable Standards	<ul style="list-style-type: none"> • CAN/CSA C22.2 No. 60601-1:14: Medical electrical equipment – Part 1: General requirements for basic safety and essential performance (Third Edition, dated March 2014) • AAMI ES60601-1: Medical electrical equipment - Part 1: General requirements for basic safety and essential performance - Consolidated reprint C1: 2009; Incorporates Amendment 2: 2010 and Amendment1:2012.
Factory/Manufacturing Location	Same as Applicant

Statement of Compliance: The product(s) identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.



Issued By: Scott Airdrie

Signature: *Scott Airdrie*

Date: December 17, 2021